

----- Forwarded message -----

From: **WebNos Governance System** <webnos@communityheartbeat.org.uk>

Date: Mon, 28 Feb 2022 at 13:26

Subject: Corfe Castle Telephone Box, Junction of East Street and Mead Road, Dorset, BH20 5EW (400061045) : Site Check submitted (3047)

To: Liz Spicer Short <rupert.liz@gmail.com>

WEBNOS

WEB NOTIFICATION SYSTEM

SITE CHECK

Corfe Castle Telephone Box, Junction of East Street and Mead Road, Dorset, BH20 5EW (400061045)

Device type:

Lifeline View (Defibtech)

Ambulance group:

South West Ambulance Service Trust

Serial number:

400061045

Ambulance Site ID :

EC0226

Date: Monday, February 28, 2022 - 13:25

Reporter: Liz Spicer-Short

Source: 92.7.192.106

SUBMITTED VALUES

Does the defibrillator have a cabinet? Yes

Is the cabinet door operational? OK - opens fine

Do all the lock buttons press okay?	Yes all OK
-------------------------------------	------------

Is there any sign of water in the cabinet or bag?	OK - clear
---	------------

Is the light working?	Yes - all OK
-----------------------	--------------

Is the thermostat set for 7+ degrees?	Everything OK
---------------------------------------	---------------

Ready indicator status	OK
------------------------	----

Is your battery inserted and has power?	Yes - All OK
---	--------------

Are the main pads connected and in date?	Yes
--	-----

Is the rescue / ready kit with the defib?	Yes - All OK
---	--------------

Frequency of checks	Weekly (recommended)
---------------------	----------------------

Cabinet status	ok
----------------	----

Copyright 2022 © Community Heartbeat Trust · Registered Charity Number 1132824